

Sulfur Compounds Analysis

1. Analytical Condition

YL 6500 Series

Oven : 80°C (5 min)-20°C/min-200 °C(4 min)

Column : DB-1(30m*0.32mm*5.0um)

Carrier gas : N₂, 1.4ml/min (SR 5:1)

Injector : Capillary 250°C

Detector : PFPD 250°C – 11.1(H₂), 14.1(Air1), 11.5(Air2)

2. Chromatogram

